

DISTILLED WATER ANALYSIS 2012

Typical Analysis

- Conductivity <math><5.0\mu\text{S}/\text{cm}</math>
- H₂O M.W. 18.01
- pH 5-8
- Chloride (as Cl) <math><5.0\text{ppm}</math>
- Nitrate <math><0.2\text{ppm}</math>
- Nitrite <math><0.003\text{ppm}</math>
- Sulphate <math><5.0\text{ppm}</math>
- Ammonia (as N) <math><0.01\text{ppm}</math>

Please note - our water is not certified for human consumption.

* Conductivity reading taken at time of manufacture. This will increase over time, due to CO₂ absorption.

www.thedistilledwatercompany.com

